



One six-station bank of the ADL Hip Simulator

ADL Hip Simulator

Specifications



ADL-Hip-12-08 specifications

General ^[1]		
Test stations	Specification	Comment
Twelve stations	See tables below	Arranged in two banks of six motion-linked stations with independent shutoff valves at each station
Load soak stations	Specification	Comment
Four stations	Active load soak	
Axial load	4500N (1000lb)	±25 mm (±1 inch)
DOF	Specification	Comment
Axial load	±4500N (±1000lb)	±25 mm (±1 inch)
Flexion	±55 degrees	20 N-m (180 in-lb)
IE rotation	±20 degrees	8 N-m (70 in-lb)
Abduction	±20 degrees	20 N-m (180 in-lb)
Actuator type	Specification	
Axial load	Servo-hydraulic	
Flexion	Servo-hydraulic	
IE rotation	Servo-hydraulic	
Abduction	Servo-hydraulic	
Control feedback	Typical method	
Axial load	Fz load cell	
Flexion	Angle sensor	
IE rotation	Angle sensor	
Abduction	Angle sensor	
Load cells	Specification	Comment
Twelve independent load cells	1 per station	
Channels	6 DOF per station	Fx, Fy, Fz, Mx, My, Mz
Type	Strain gage	Amplifier included in controls
Physical specifications	Specification	
Length	208 cm (82 in)	
Width	100 cm (39 in)	
Height	178 cm (70 in)	
Weight	1100 kg (2500 lb)	
Hydraulic system (quoted separately)	Specification	Comment
Type	External HPU	Required
Pressure	800 psi	Required
Required flow	15 GPM	Required
Oil temperature	38 °C	Recommended temperature at inlet
Power requirement ^[2]	Specification	Comment
Electric	115 VAC, 20 Amp	1 phase, 50/60 Hz

Dynamic performance			
Item	Maximum repetition rate (repetitions/second) ^[3]	Maximum spectral content (Hz) ^[4]	
Controller	30 Hz	30 Hz	
DOF	Typical repetition rate (repetitions/second) ^[3]	Maximum spectral content (Hz) ^[4]	RMS error (% FS) ^[5]
Axial load	2.0 Hz	10 Hz	< 1 %
Flexion extension	2.0 Hz	10 Hz	< 1 %
IE rotation	2.0 Hz	10 Hz	< 1 %
Abduction	2.0 Hz	10 Hz	< 1 %



ADL-Hip-12-08 specifications

Measurement instrumentation			
Data acquisition	Channel	Range	Comment
Data rate	All channels	10-1000 samples/sec	User selection
ADC Resolution	All channels	16 bit	
Digital filters	All channels	10-500Hz	User selection
Anti-aliasing filters	All channels	360 Hz	
Strain gage conditioning	Channel	Range	Comment
Gain	All channels	1000, 2000, 4000	Jumper selection
Excitation	All channels	10 Volt	
Multi-axis load cell – hip	Channel	Range	Sensitivity
Axial load	Fz	4400 N	1.5 $\mu\text{V}/\text{V}\bullet\text{N}$
AP force	Fy	500 N	6.0 $\mu\text{V}/\text{V}\bullet\text{N}$
ML force	Fx	500 N	6.0 $\mu\text{V}/\text{V}\bullet\text{N}$
Flexion moment	Mx	25 Nm	70.8 $\mu\text{V}/\text{V}\bullet\text{Nm}$
Abduction moment	My	25 Nm	70.8 $\mu\text{V}/\text{V}\bullet\text{Nm}$
Axial moment	Mz	25 Nm	53.1 $\mu\text{V}/\text{V}\bullet\text{Nm}$
Angle and position	Channel	Range	~Resolution
Axial position	VP	38 mm	0.1 mm
Flexion	Flex	$\pm 55^\circ$	0.1°
IE rotation	IE	$\pm 20^\circ$	0.1°
Abduction	AD	$\pm 20^\circ$	0.1°
Other sensors	Channel	Comment	
Serum temperature			
Oil temperature			
Fluid level	High/Low	Safety shutoff if leak detected	
Hydraulic pressure			
Cal amp input			

Environmental conditioning for specimens		
Specimen fluid recirculation	Specification	Comment
Pump	100 ml/min	60 RPM peristaltic pump with #25 silicone tubing
Reservoir	500 ml	Stainless steel tank
Fluid level	High/Low	Magnetic sensor/float
Specimen fluid temperature	Specification	Comment
Temperature controller		Heater/chiller
Chilling	500 watts	
Heating	800 watts	
Power		115 volts, 12 amps
Specimen fluid	Specification	
Suitable fluids	Bovine serum, saline solution, water	



ADL-Hip-12-08 specifications

Control system			
NetControl interface	Specification	Comments	
Supplied computer hardware	One set per bank	Windows XP PC, monitor, keyboard, mouse	
Ethernet connection	10 mbps	Coaxial cable	
Real-time controller	Channels	Range	Comments
DSP controller		AD 2181	
Update rate		2000 Hz	
Output channels	16	± 10 volt	
Input channels	72	Bridge inputs	Strain gage
Analog inputs	30	± 10 volt	High-level analog
Digital inputs	8	TTL	Digital I/O
Control modes	Channels		
PID	16		
Gain scheduling	16		
Adaptive Control	16		
Virtual Soft Tissue	8		
Nested loops	8		
Waveform generator	Channels	Range	Comments
Channels	8-16		
Repetition rate		0.01 to 30 Hz	
Programmable		256 points	Interpolated
Event monitor	Channels	Specification	Comments
Threshold trigger	16		Rising or falling edge
Response time	All channels	0.0005 seconds	
Programmable response	All channels		Soft stop, hold, shut down
Digital outputs	Channels	Update rate	
Reference waveforms	8-16	2000 Hz	
Servo drive signals	8-16	2000 Hz	
Soft tissue constraint	8	2000 Hz	
Sum signals	8	2000 Hz	
Digital loop filters	Channels	Update rate	
	16	30-1000 Hz	

[1] Specifications may change without notice.

[2] The system is normally delivered configured for the indicated power requirements. If your available power differs in phase or voltage, please contact AMTI. The system requires cooling water for operation. This is usually available from your laboratory's infrastructure – if not available, please contact AMTI for additional information on manufacturers of suitable chillers.

[3] The repetition rate corresponds to the maximum rate at which satisfactory performance will be achieved running the ISO standard gait cycle waveforms for hip testing. This is a somewhat subjective indication of dynamic performance. Typically, overall tracking performance is reduced with higher frequency of operation.

[4] The ISO waveforms contain spectral content in considerable excess of the fundamental driving frequency. Analysis of these waveforms indicates that tracking performance at a 1 Hz repetition rate is excellent up to the indicated frequency.

[5] The RMS error provides a measure of the simulator's tracking performance (the extent that the machine's outputs differ from the target inputs). These values are typical for testing at a 1 Hz repetition rate while running the ISO waveforms and represent standard results while evaluating conventional prosthetics using AMTI's Adaptive Control Technology (iterative learning control algorithm). Different prosthetic devices or conditions may result in an increased or decreased tracking error.